



Ref. Certif. No.

DK-37597-M1-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

Product
ProduitName and address of the applicant
Nom et adresse du demandeurName and address of the manufacturer
Nom et adresse du fabricantName and address of the factory
Nom et adresse de l'usineNote: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} pageRatings and principal characteristics
Valeurs nominales et caractéristiques principalesTrademark (if any)
Marque de fabrique (si elle existe)
Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeurModel / Type Ref.
Ref. De typeAdditional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{ème} pageA sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à laAs shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

CERTIFICAT D'ESSAI OC

AC/DC Power Module

TRACO ELECTRONIC AG
SIHLBRUGGSTRASSE 111
CH-6340 BAAR, SWITZERLANDTRACO ELECTRONIC AG
SIHLBRUGGSTRASSE 111
CH-6340 BAAR SWITZERLAND☐ Additional Information on page 2
See Page 2TRACO POWER
WMTTMM 24aX, TMM 24bX
See Page 2☒ Additional Information on page 2

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

E188913-MED27 issued on 2014-03-17

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

- ☐ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☒ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnamesDate: 2014-03-18
Original Issue Date: 2014-03-17

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-37597-M1-UL

Model Details:

TMM 24aX, TMM 24bX (where a can be 105, 109, 112, 115, 124; b can be 212, 215; X can be C or blank)

Ratings:

Input: 100-240Vac, 60/50 Hz, 0.6A- 0.3A

Output:

Models Output

TMM 24105X 5.0 V dc, 3.0 A

TMM 24109X 9.0 V dc, 2.666 A

TMM 24112X 12 V dc, 2.0 A

TMM 24115X 15 V dc, 1.6 A

TMM 24124X 24 V dc, 1.0 A

TMM 24212X +/-12 V dc, +/-1.0 A

TMM 24215X +/-15 V dc, +/-0.8 A

Additional Information:

Additionally evaluated to EN 60601-1: 2006 /A1:2013; National Differences specified in the CB Test Report.

The risk management requirements of the standard were not addressed

Correction to Certificate only: to correct report issue date.

Additional information (if necessary)

Information complémentaire (si nécessaire)



☐ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

☒ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2014-03-18

Original Issue Date: 2014-03-17

Signature:

Jan-Erik Storgaard