

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**CB TEST CERTIFICATE**

Product

Switch Mode Power Supply for building-in (DIN RAIL)

Name and address of the applicant

Traco Power Solutions Ltd.
Whitemill Industrial Estate Wexford, White Mill Road,
Y35 YH66, Ireland

Name and address of the manufacturer

Traco Power Solutions Ltd.
Whitemill Industrial Estate Wexford, White Mill Road,
Y35 YH66, Ireland

Name and address of the factory

Traco Power Solutions Ltd.
Whitemill Industrial Estate Wexford, White Mill Road,
Y35 YH66, Ireland*Note: When more than one factory, please report on page 2*☐ Additional Information on page 2

Ratings and principal characteristics

See page 2

Trademark / Brand (if any)

TRACO POWER

Customer's Testing Facility (CTF) Stage used

/

Model / Type Ref.

TSP-WR series (see page 2)

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to EN 61010-1:2010 + A1:2019 and EN IEC 61010-2-201:2018; National Differences specified in the CB Test Report.

☒ Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010 + A1:2016
IEC 61010-2-201: 2017

As shown in the Test Report Ref. No. which forms part of this Certificate

T223-0744/20, T223-0745/20 (2020-11-24)

This CB Test Certificate is issued by the National Certification Body

SIQ Ljubljana, Mašera-Spasičeva ulica 10, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

SIQ Ljubljana is accredited by Slovenian Accreditation with accreditation number CP-001 in the field of certification of products, processes and services.

Date: 2020-11-24

Signature: Bojan Pečavar



Model / Type Ref.:

TRACO POWER Model reference	Weidmüller Interface GmbH Co.KG Model reference	Manufacturer Model reference
TSP180-124WRxxxx	CP T SNT2 180W 24V 7,5A	120PSP184xxxx
TSP360-124WRxxxx	CP T SNT2 360W 24V 15A	240PSP184xxxx
TSP600-124WRxxxx	CP T SNT2 600W 24V 25A	480PSP184xxxx

* Each model can be alternatively marked with a suffix xxxx for non-safety relevant changes. No impact on safety.

Ratings and principal characteristics:
Input:

TSP180-124WR: 115/230/400-500 Vac; 4,0/2,0/1,0 A; 50/60 Hz

TSP360-124WR: 115/230/400-500 Vac; 6,0/3,0/1,5 A; 50/60 Hz

TSP600-124WR: 115/230/400-500 Vac; 10,0/5,0/3,0 A; 50/60 Hz

Output:

TSP180-124WR: 24 Vdc; 7,5 A

TSP360-124WR: 24 Vdc; 15,0 A

TSP600-124WR: 24 Vdc; 25,0 A

Outputs are adjustable from 24-28 Vdc

Additional information (if necessary)

Date: 2020-11-24

Signature: Bojan Pečavar

