



Ref. Certif. No.

**US-19491-UL**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE**Product  
ProduitName and address of the applicant  
Nom et adresse du demandeurName and address of the manufacturer  
Nom et adresse du fabricantName and address of the factory  
Nom et adresse de l'usineNote: When more than one factory, please report on page 2  
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2<sup>ème</sup> pageRatings and principal characteristics  
Valeurs nominales et caractéristiques principalesTrademark (if any)  
Marque de fabrique (si elle existe)Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeurModel / Type Ref.  
Ref. De typeAdditional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>ème</sup> pageA sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à laAs shown in the Test Report Ref. No. which forms  
part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat**CERTIFICAT D'ESSAI OC**

Built-in type Power Supply

TRACO ELECTRONIC AG  
JENATSCHSTRASSE 1  
ZURICH, CH-8002 Switzerland☒ Additional Information on page 2  
See Page 2

TRACO POWER

TMT30105, TMT30105C, TMT30112, TMT30112C, TMT30115,  
TMT30115C, See Page 2☒ Additional Information on page 2

IEC 60601-1(ed.3)

E141988-A66-CB-1 issued on 2012-08-03

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme National de Certification

- ☒ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2012-08-03

Signature:

  
Jolanta M. Wroblewska



Ref. Certif. No.

**US-19491-UL**

**Model Details:**

TMT30124, TMT30124C, TMT30212, TMT30212C, TMT30215, TMT30215C

**Factories:**

**Ratings:**

For Models TMT30105, TMT30105C

Input: 115-230 V~, 50/60 Hz, 0.56-0.35 A,

Output: 5 Vdc, 6000 mA

For Models TMT30112, TMT30112C

Input: 115-230 V~, 50/60 Hz, 0.56-0.35 A,

Output: 12 Vdc, 2500 mA

For Models TMT30115, TMT30115C

Input: 115-230 V~, 50/60 Hz, 0.56-0.35 A,

Output: 15 Vdc, 2000mA

For Models TMT30124, TMT30124C

Input: 115-230 V~, 50/60 Hz, 0.56-0.35 A,

Output: 24 Vdc, 1250 mA

For Models TMT30212, TMT30212C

Input: 115-230 V~, 50/60 Hz, 0.57-0.36 A,

Output:  $\pm 12$  Vdc, 1250 mA

For Models TMT30215, TMT30215C

Input: 115-230 V~, 50/60 Hz, 0.57-0.36 A,

Output:  $\pm 15$  Vdc, 1000 mA

**Additional Information:**

Risk Management was not included in this investigation.

Additionally evaluated to EN 60601-1:2006; National Differences specified in the CB Test Report. Build-in component.

**Additional information (if necessary)**

**Information complémentaire (si nécessaire)**



☒ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

☐ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2012-08-03

Signature:

Jolanta M. Wroblewska