

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

Product
Produit

Name and address of the applicant
Nom et adresse du demandeur

Name and address of the manufacturer
Nom et adresse du fabricant

Name and address of the factory
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Trademark (if any)
Marque de fabrique (si elle existe)
Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

Model / Type Ref.
Ref. De type

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{ème} page

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

CERTIFICAT D'ESSAI OC

Built-in type Power Supply

TRACO ELECTRONIC AG
SIHLBRUGGSTRASSE 111, CH-6340 BAAR, SWITZERLAND

☐ Additional Information on page 2
See Page 2

TRACO POWER

TMT50105, TMT50112, TMT50115, TMT50124, TMT50148,
TMT50105C, TMT50112C, TMT50115C, TMT50124C,
TMT50148C

☒ Additional Information on page 2

IEC 60950-1(ed.2), IEC 60950-1(ed.2);am1,
IEC 60950-1(ed.2);am2

E141988-A41-CB-3 issued on 2015-08-31

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



- ☐ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☐ UL (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
☒ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2015-08-31

Signature:



Yoshihiro Yamamoto

Ratings:

For Models TMT50105, TMT50105C
Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A,
Output: 5 Vdc, 9000 mA

For Models TMT50112, TMT50112C
Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A,
Output: 12 Vdc, 4200 mA

For Models TMT50115, TMT50115C
Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A
Output: 15 Vdc, 3400mA

For Models TMT50124, TMT50124C
Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A
Output: 24 Vdc, 2300 mA

For Models TMT50148, TMT50148C
Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A
Output: 48 Vdc, 1150 mA

Additional Information:

Additionally evaluated to EN 60950-1:2006/A11:2009/A1:2010/A12:2011/A2:2013
National Differences specified in the CB Test Report.

Additional information (if necessary)

Information complémentaire (si nécessaire)



☐ UL (US) 333 Pfingsten Rd IL 60062 Northbrook USA

☐ UL (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK

☒ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

☐ UL (CA) 7 Underwriters Road Toronto M1R 3B4 Ontario CANADA

For full legal entity names see www.ul.com/nbnames

Date: 2015-08-31

Signature:

Yoshihiro Yamamoto