

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

## CB TEST CERTIFICATE

Product  
Produit

Name and address of the applicant  
Nom et adresse du demandeur

Name and address of the manufacturer  
Nom et adresse du fabricant

Name and address of the factory  
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

Trademark (if any)  
Marque de fabrique (si elle existe)  
Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais constructeur

Model / Type Ref.  
Ref. De type

Additional information (if necessary may also be reported on page 2)  
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2<sup>ème</sup> page

A sample of the product was tested and found to be in conformity with  
Un échantillon de ce produit a été essayé et a été considéré conforme à la

As shown in the Test Report Ref. No. which forms part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

## CERTIFICAT D'ESSAI OC

AC/DC Power Module

TRACO ELECTRONIC AG  
SIHLBRUGGSTRASSE 111  
CH-6340 BAAR, SWITZERLAND

TRACO ELECTRONIC AG  
SIHLBRUGGSTRASSE 111  
CH-6340 BAAR, SWITZERLAND

☐ Additional Information on page 2  
See Page 2

TRACO POWER

WMT

TMM 40aX, TMM 40bX series  
See Page 2

Additionally evaluated to EN 60601-1:2006/A1:2013.  
Risk Management was not included in this investigation.

☐ Additional Information on page 2

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

E188913-MED32 issued on 2014-03-20

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme National de Certification



- ☐ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☒ UL (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2014-03-28

Signature:

Jan-Erik Storgaard

## Model Details:

TMM 40aX, TMM 40bX series (where a may be 105, 112, 115 or 124; where b may be 212 or 215; X can be C or blank)

## Ratings:

Input: 100-240Vac, 60/50 Hz, 1.1A-0.6A

## Output:

Models                  Output

TMM 40105X    5.0 V dc, 8.0A

TMM 40112X    12 V dc, 3.33A

TMM 40115X    15 V dc, 2.66A

TMM 40124X    24 V dc, 1.66A

TMM 40212X    +/-12 V dc, +/-1.66 A

TMM 40215X    +/-15 V dc, +/-1.33 A

## Additional information (if necessary)

## Information complémentaire (si nécessaire)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2014-03-28

Signature:

Jan-Erik Storgaard