

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Power Supply

Name and address of the applicant

Traco Power Solutions Ltd.
Whitemill Industrial Estate Wexford, Whitemill Road,
Wexford Y35 YH66, Ireland

Name and address of the manufacturer

Traco Power Solutions Ltd.
Whitemill Industrial Estate Wexford, Whitemill Road,
Wexford Y35 YH66, Ireland

Name and address of the factory

Traco Power Solutions Ltd.
Whitemill Industrial Estate Wexford, Whitemill Road,
Wexford Y35 YH66, Ireland

Note: When more than one factory, please report on page 2
☐ Additional Information on page 2

Ratings and principal characteristics

See page 2

Trademark / Brand (if any)

TRACO POWER

Customer's Testing Facility (CTF) Stage used

/

Model / Type Ref.

TEX series (see page 2)

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to EN 61010-1:2010 + A1:2019 and EN IEC 61010-2-201:2018; National Differences specified in the CB Test Report.

☒ Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010 + A1:2016
IEC 61010-2-201:2017

As shown in the Test Report Ref. No. which forms part of this Certificate

T223-0096/21, T223-0097/21 (2021-03-04)


This CB Test Certificate is issued by the National Certification Body


SIQ Ljubljana, Mašera-Spasičeva ulica 10, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

SIQ Ljubljana is accredited by Slovenian Accreditation with accreditation number CP-001 in the field of certification of products, processes and services.

Date: 2021-03-04

Signature: Bojan Pečavar



Model / Type Ref.:

TRACO POWER Model reference	Manufacturer Model reference
TEX120-124	120WPP182
TEX120-124	120WPP184
The TEX120-112 is also available as an alternative model marked with internal model reference T1075-01A	

Ratings and principal characteristics:

TEX120-112

In: 100-240 Vac; 2,2-1,0 A; 50/60 Hz

Out: 12 Vdc / 8 A

TEX120-124

In: 100-240 Vac; 2,4-1,2 A; 50/60 Hz

Out: 24 Vdc / 5 A

Additional information (if necessary)

Date: 2021-03-04

Signature: Bojan Pečavar

