



# Certificate of Compliance

**Certificate:** 2655908

**Master Contract:** 219759

**Project:** 2655908

**Date Issued:** September 20, 2013

**Issued to:** Convertec Limited

Whitemill Industrial Estate  
Wexford, Wexford  
Ireland  
Attention: Michael Carey

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Patric Charbonneau*

**Issued by:** Patric Charbonneau, P.T.

## **PRODUCTS**

**CLASS 3631 06** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 86** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Switching Mode Power Supplies, Trademarks: TRACO POWER or Convertec Power Electronics, models and ratings as follows:

TRACO POWER	Convertec Power Electronics	Manufacturer Reference	Ratings
TEX 120-112	IPEX 012.080	120WPP182	Input: 100-240Vac, 50/60Hz, 2.2-1.0A Output: 12Vdc, 8A
TEX 120-124	IPEX 024.050	120WPP184	Input: 100-240Vac, 50/60Hz, 2.4-1.2A Output: 24Vdc, 5A



**Certificate:** 2655908

**Master Contract:** 219759

**Project:** 2655908

**Date Issued:** September 20, 2013

---

Notes:

- The above models are cord-connected (using special type connectors), Equipment Class I, Pollution Degree 2, Installation Category II.
- Mode of operation: Continuous.
- Environmental Conditions: Extended: -40°C to +85°C (with output power de-rating of 2%/K above 60°C), maximum altitude of 3000 m, up to 100% relative humidity with condensation.
- The units have been evaluated for use in ordinary dry locations only.

**APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-12 – Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

ANSI/UL 61010-1, Third Edition – Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements