

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

## CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product  
Produit

Built-in type Power Supply

Name and address of the applicant  
Nom et adresse du demandeur

Name and address of the manufacturer  
Nom et adresse du fabricant

TRACO ELECTRONIC AG  
JENATSCHSTRASSE 1  
CH-8002 ZURICH, SWITZERLAND

Name and address of the factory  
Nom et adresse de l'usine

*Note: When more than one factory, please report on page 2*  
*Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page*

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

☒ Additional Information on page 2  
See Page 2

Trademark (if any)  
Marque de fabrique (si elle existe)

TRACO POWER

Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais constructeur

Model / Type Ref.  
Ref. De type

TMT50105, TMT50105C, TMT50112, TMT50112C, TMT50115,  
TMT50115C, See page 2

Additional information (if necessary may also be reported on page 2)  
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2<sup>ème</sup> page

Additionally evaluated to EN 60601-1: 2006; National Differences specified in the CB Test Report.

☐ Additional Information on page 2

A sample of the product was tested and found to be in conformity with  
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

E141988-A67-CB-1 issued on 2012-08-03

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme National de Certification




- ☒ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2012-08-03

Signature:



Jolanta M. Wroblewska



Ref. Certif. No.

**US-19496-UL**

**Model Details:**

TMT50124, TMT50124C, TMT50148, TMT50148C

**Factories:**

**Ratings:**

For Models TMT50105, TMT50105C

Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A

Output: 5 Vdc, 9000 mA

For Models TMT50112, TMT50112C

Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A

Output: 12 Vdc, 4200 mA

For Models TMT50115, TMT50115C

Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A

Output: 15 Vdc, 3400mA

For Models TMT50124, TMT50124C

Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A

Output: 24 Vdc, 2300 mA

For Models TMT50148, TMT50148C

Input: 115-230 V~, 50/60 Hz, 1.0-0.6 A

Output: 48 Vdc, 1150 mA

**Additional information (if necessary)**

**Information complémentaire (si nécessaire)**



☒ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

☐ UL (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK

☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2012-08-03

Signature:

Jolanta M. Wroblewska