



Ref. Certif. No.

SI-7773

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Power supply for building-in

Name and address of the applicant

Traco Power Solutions Ltd.  
Whitemill Industrial Estate, Whitemill Road, Wexford  
Y35 YH66, Ireland

Name and address of the manufacturer

Traco Power Solutions Ltd.  
Whitemill Industrial Estate, Whitemill Road, Wexford  
Y35 YH66, Ireland

Name and address of the factory

Traco Power Solutions Ltd.  
Whitemill Industrial Estate, Whitemill Road, Wexford  
Y35 YH66, Ireland

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

24 W models: Input: 100-240 Vac; 50/60 Hz; 0,5-0,25 A  
36 W models: Input: 100-240 Vac; 50/60 Hz; 0,7-0,4 A  
See page 2

Trademark (if any)

**TRACO<sup>®</sup>  
POWER**

Customer's Testing Facility (CTF) Stage used

/

Model / Type Ref.

24 W models: TMW 24-105, TMW 24-112, TMW 24-124  
36 W models: TMW 36-112; TMW 36-124  
See page 2

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed.

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005 + A1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

T223-0049/20, issued on 2020-01-27

This CB Test Certificate is issued by the National Certification Body



SIQ Ljubljana, Tržaška cesta 2, SI-1000 Ljubljana, Slovenia  
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

SIQ Ljubljana is accredited by Slovenian Accreditation with accreditation number CP-001 in the field of certification of products, processes and services.

Date: 2020-01-27

Signature: Bojan Pečavar

**Ratings and principal characteristics:****24 W models:**

Input: 100-240 Vac; 50/60 Hz; 0,5-0,25 A

Output:

TMW 24-105: 5,1 Vdc; 4,0 A

TMW 24-112: 12 Vdc; 2,0 A

TMW 24-124: 24 Vdc; 1,0 A

**36 W models:**

Input: 100-240 Vac; 50/60 Hz; 0,7-0,4 A

Output:

TMW 36-112: 12 Vdc; 3,0 A

TMW 36-124: 24 Vdc; 1,5 A

**Model / Type Ref.:****24 W models:** TMW 24-105, TMW 24-112, TMW 24-124**36 W models:** TMW 36-112; TMW 36-124

\* Models can be additionally marked with "P", where "P" indicates pin version

\*\* models can be additionally marked with "xxaaaaa", where "x" or "a" can be any alphanumeric character, blank or dash (not impact on safety)

**Additional information:**

Power supply for building-in also complies with the following standards:

- IEC 60335-1:2010 + A1:2013 + A2:2013
- IEC 61558-2-16:2009 + A1:2013
- IEC 61558-2-8:2010
- IEC 61347-2-13:2014 + A1:2016
- IEC 62368-1:2014

**Additional information (if necessary)**

This CB Test Certificate substitutes previously issued CB Test Certificate No. SI-7437, issued on 2019-11-07, due to modification of the test report.

Date: **Error! Reference source not found.**

Signature: Bojan Pečavar

