

DECLARATION OF CONFORMITY



AC/DC Industrial Power Supply, TCL 120 Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TCL 120-112(z), TCL 120-112C(z), TCL 120-124(z), TCL 120-124C(z)**
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks


Directives: **LVD: 2014/35/EU
EMC: 2014/30/EU
RoHS: 2011/65/EU**

Safety standards: **EN 61010-1:2010 + A1:2019
EN 61010-2-201:2018**

EMC standards: **EN 61000-6-2:2005 + AC:2005
EN 61000-6-3:2007 + A1:2011 + AC:2012
EN 55032:2012**

Year of CE marking: **2003**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.



Stefan Schaffhauser, CEO

DECLARATION OF CONFORMITY



AC/DC Industrial Power Supply, TCL 120 Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TCL 120-112(z), TCL 120-112C(z), TCL 120-124(z), TCL 120-124C(z)**
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks


Regulations: **Electrical Equipment (Safety) Regulations 2016
Electromagnetic Compatibility Regulations 2016
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012**

Safety standards: **EN 61010-1:2010 + A1:2019
EN 61010-2-201:2018**

EMC standards: **EN 61000-6-2:2005 + AC:2005
EN 61000-6-3:2007 + A1:2011 + AC:2012
EN 55032:2012**

Year of UKCA marking: **2021**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.



Stefan Schaffhauser, CEO