

DECLARATION OF CONFORMITY



Function Module, TIB-BCMU Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TIB 240-124BCMU(z)**
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Directives: **LVD: 2014/35/EU**
EMC: 2014/30/EU
RoHS: 2011/65/EU

Safety standards: **EN 62368-1:2014 + A11:2017**
EN 61010-1:2010 + A1:2019
EN IEC 61010-2-201:2018

EMC standards: **EN 61000-6-3:2007 + A1:2011** (only when used in correspondence with a TIB power supply)
EN 62040-2:2006 (only when used in correspondence with a TIB power supply)

Year of CE marking: **2022**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.



Stefan Schaffhauser, CEO

Baar, February 28, 2024

DECLARATION OF CONFORMITY



Function Module, TIB-BCMU Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TIB 240-124BCMU(z)**
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks


Regulations: **Electrical Equipment (Safety) Regulations 2016**
Electromagnetic Compatibility Regulations 2016
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

Safety standards: **EN 62368-1:2014 + A11:2017**
EN 61010-1:2010 + A1:2019
EN IEC 61010-2-201:2018

EMC standards: **EN 61000-6-3:2007 + A1:2011** (only when used in correspondence with a TIB power supply)
EN 62040-2:2006 (only when used in correspondence with a TIB power supply)

Year of UKCA marking: **2022**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.



Stefan Schaffhauser, CEO

Baar, February 28, 2024