

## DECLARATION OF CONFORMITY



### AC/DC Medical Power Supply, TMW 24P Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TMW 24-105P(z), TMW 24-112P(z), TMW 24-124P(z)**  
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Directives: **LVD: 2014/35/EU  
EMC: 2014/30/EU  
RoHS: 2011/65/EU**

Safety standards: **EN 62368-1:2014 + A11:2017**

EMC standards: **EN 61000-6-2:2005 + AC:2005  
EN 61000-6-3:2007 + A1:2011 + AC:2012  
EN 55032:2012  
EN 61000-3-2:2014**

Year of CE marking: **2021**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.



Stefan Schaffhauser, CEO

Baar, February 21, 2024

## DECLARATION OF CONFORMITY



### AC/DC Medical Power Supply, TMW 24P Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TMW 24-105P(z), TMW 24-112P(z), TMW 24-124P(z)**  
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Regulations: **Electrical Equipment (Safety) Regulations 2016  
Electromagnetic Compatibility Regulations 2016  
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012**

Safety standards: **EN 62368-1:2014 + A11:2017**

EMC standards: **EN 61000-6-2:2005 + AC:2005  
EN 61000-6-3:2007 + A1:2011 + AC:2012  
EN 55032:2012  
EN 61000-3-2:2014**

Year of UKCA marking: **2021**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.



Stefan Schaffhauser, CEO

Baar, February 21, 2024