

## DECLARATION OF CONFORMITY



### AC/DC Power Supply, TPI 180-M Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TPI 180-112-M(z), TPI 180-115-M(z), TPI 180-118-M(z), TPI 180-124-M(z), TPI 180-128-M(z), TPI 180-136-M(z), TPI 180-148-M(z), TPI 180-153-M(z)**  
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Directives: **LVD: 2014/35/EU  
EMC: 2014/30/EU  
RoHS: 2011/65/EU**

Safety standards: **EN 62368-1:2014 + A11:2017**

EMC standards: **EN 55032:2015 + AC:2016  
EN 55035:2017  
EN 61000-3-2:2014  
EN 61000-3-3:2013**

Year of CE marking: **2021**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.



Stefan Schaffhauser, CEO

Baar, February 21, 2024

## DECLARATION OF CONFORMITY



### AC/DC Power Supply, TPI 180-M Series

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TPI 180-112-M(z), TPI 180-115-M(z), TPI 180-118-M(z), TPI 180-124-M(z), TPI 180-128-M(z), TPI 180-136-M(z), TPI 180-148-M(z), TPI 180-153-M(z)**  
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Regulations: **Electrical Equipment (Safety) Regulations 2016  
Electromagnetic Compatibility Regulations 2016  
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012**

Safety standards: **EN 62368-1:2014 + A11:2017**

EMC standards: **EN 55032:2015 + AC:2016  
EN 55035:2017  
EN 61000-3-2:2014  
EN 61000-3-3:2013**

Year of UKCA marking: **2021**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.



Stefan Schaffhauser, CEO

Baar, February 21, 2024