

DECLARATION OF CONFORMITY**AC/DC Medical Power Supply, TPP 30-D Series**

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TPP 30-103-D(z), TPP 30-105-D(z), TPP 30-109-D(z), TPP 30-112-D(z), TPP 30-115-D(z), TPP 30-124-D(z), TPP 30-136-D(z), TPP 30-148-D(z)**

(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Directives: **LVD: 2014/35/EU
EMC: 2014/30/EU
RoHS: 2011/65/EU**

Safety standards: **EN 62368-1:2014 + A11:2017
EN 60601-1:2006 + A1:2013**

EMC standards: **EN 55032:2015 + A11:2020
EN 61000-3-2:2019 + A1:2021
EN 61000-3-3:2013 + A1:2019 + A2:2021 + AC:2022-01
EN 55035:2017 + A11:2020
EN 60601-1-2:2015**

Year of CE marking: **2017**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.

Stefan Schaffhauser, CEO

Baar, February 21, 2024

DECLARATION OF CONFORMITY**AC/DC Medical Power Supply, TPP 30-D Series**

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TPP 30-103-D(z), TPP 30-105-D(z), TPP 30-109-D(z), TPP 30-112-D(z), TPP 30-115-D(z), TPP 30-124-D(z), TPP 30-136-D(z), TPP 30-148-D(z)**

(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Regulations: **Electrical Equipment (Safety) Regulations 2016
Electromagnetic Compatibility Regulations 2016
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012**

Safety standards: **EN 62368-1:2014 + A11:2017
EN 60601-1:2006 + A1:2013**

EMC standards: **EN 55032:2015 + A11:2020
EN 61000-3-2:2019 + A1:2021
EN 61000-3-3:2013 + A1:2019 + A2:2021 + AC:2022-01
EN 55035:2017 + A11:2020
EN 60601-1-2:2015**

Year of UKCA marking: **2021**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.

Stefan Schaffhauser, CEO

Baar, February 21, 2024