

DECLARATION OF CONFORMITY**AC/DC Medical Power Supply, TPP 40 Series**

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TPP 40-105(z), TPP 40-112(z), TPP 40-115(z), TPP 40-124(z), TPP 40-221(z), TPP 40-231(z), TPP 40-251(z), TPP 40-321M2(z), TPP 40-331M3(z), TPP 40-3512(z)**
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Directives: **LVD: 2014/35/EU
EMC: 2014/30/EU
RoHS: 2011/65/EU**

Safety standards: **EN 62368-1:2014 + A11:2017
EN 60601-1:2006 + A1:2013**

EMC standards: **EN 55011:2009 + A1:2010
EN 55032:2012
EN 55024:2010 + A1:2010
EN 60601-1-2:2006 + A1:2013 + A12:2014**

Year of CE marking: **2013**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.

Stefan Schaffhauser, CEO

Baar, February 21, 2024

DECLARATION OF CONFORMITY**AC/DC Medical Power Supply, TPP 40 Series**

Traco Electronic Co. Ltd. declares under its sole responsibility that the product(s) to which this declaration relates is in conformity with the following directives, standard(s) and other normative document(s).

List of models: **TPP 40-105(z), TPP 40-112(z), TPP 40-115(z), TPP 40-124(z), TPP 40-221(z), TPP 40-231(z), TPP 40-251(z), TPP 40-321M2(z), TPP 40-331M3(z), TPP 40-3512(z)**
(z) can be blank, any number of alphanumeric characters, "-", or punctuation marks

Regulations: **Electrical Equipment (Safety) Regulations 2016
Electromagnetic Compatibility Regulations 2016
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012**

Safety standards: **EN 62368-1:2014 + A11:2017
EN 60601-1:2006 + A1:2013**

EMC standards: **EN 55011:2009 + A1:2010
EN 55032:2012
EN 55024:2010 + A1:2010
EN 60601-1-2:2006 + A1:2013 + A12:2014**

Year of UKCA marking: **2021**

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.

Stefan Schaffhauser, CEO

Baar, February 21, 2024